

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040825\
 Data File : VX045645.D
 Acq On : 08 Apr 2025 12:28
 Operator : JC/MD
 Sample : Q1737-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RT3069

Quant Time: Apr 09 01:36:31 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	60517	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	117735	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	112767	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	47540	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	61747	55.794	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 111.580%	
35) Dibromofluoromethane	5.379	113	43858	52.504	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 105.000%	
50) Toluene-d8	8.647	98	152623	52.346	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 104.700%	
62) 4-Bromofluorobenzene	11.079	95	56901	53.578	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 107.160%	
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	104893	230.817	ug/l	96
18) Methyl Acetate	2.709	43	2032	1.942	ug/l	96
20) Methylene Chloride	2.782	84	8550	10.050	ug/l	93
25) 2-Butanone	4.568	43	21068	31.671	ug/l	99
40) Benzene	6.031	78	38671	11.227	ug/l	97
52) Toluene	8.720	92	2758	1.323	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040825\
Data File : VX045645.D
Acq On : 08 Apr 2025 12:28
Operator : JC/MD
Sample : Q1737-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RT3069

Quant Time: Apr 09 01:36:31 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

