

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040825\
 Data File : VX045647.D
 Acq On : 08 Apr 2025 13:14
 Operator : JC/MD
 Sample : PB167492ZHE05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB167492ZHE05

Quant Time: Apr 09 01:37:31 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	65890	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	133175	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	127373	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	51216	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	66836	55.468	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	110.940%
35) Dibromofluoromethane	5.385	113	48062	50.866	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.740%
50) Toluene-d8	8.647	98	169300	51.334	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	102.660%
62) 4-Bromofluorobenzene	11.079	95	63467	52.832	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	105.660%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	109437	221.179	ug/l	97
18) Methyl Acetate	2.709	43	2365	2.076	ug/l	95
20) Methylene Chloride	2.782	84	11995	12.949	ug/l	91
25) 2-Butanone	4.568	43	13017	17.973	ug/l #	84
43) Isopropyl Acetate	6.348	43	9024	3.703	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040825\
Data File : VX045647.D
Acq On : 08 Apr 2025 13:14
Operator : JC/MD
Sample : PB167492ZHE05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB167492ZHE05

Quant Time: Apr 09 01:37:31 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

