

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040825\
 Data File : VX045654.D
 Acq On : 08 Apr 2025 15:58
 Operator : JC/MD
 Sample : PB167492ZHE03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB167492ZHE03

Quant Time: Apr 09 01:40:53 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	67363	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	132718	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	124907	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	51698	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	67476	54.774	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 109.540%			
35) Dibromofluoromethane	5.379	113	48421	51.422	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 102.840%			
50) Toluene-d8	8.647	98	170639	51.918	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 103.840%			
62) 4-Bromofluorobenzene	11.079	95	61958	51.754	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 103.500%			
Target Compounds					Qvalue	
16) Acetone	2.386	43	107737	212.982	ug/l	98
18) Methyl Acetate	2.703	43	2581	2.216	ug/l	96
20) Methylene Chloride	2.782	84	12090	12.766	ug/l	97
25) 2-Butanone	4.568	43	11604	15.671	ug/l	98
43) Isopropyl Acetate	6.342	43	8608	3.545	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040825\
Data File : VX045654.D
Acq On : 08 Apr 2025 15:58
Operator : JC/MD
Sample : PB167492ZHE03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB167492ZHE03

Quant Time: Apr 09 01:40:53 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

