

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040825\
 Data File : VX045655.D
 Acq On : 08 Apr 2025 16:21
 Operator : JC/MD
 Sample : Q1745-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IB-6A-WC

Quant Time: Apr 09 01:41:21 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	61215	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	122965	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	114251	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	46512	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	63067	56.337	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	112.680%
35) Dibromofluoromethane	5.385	113	44521	51.031	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.060%
50) Toluene-d8	8.647	98	154694	50.800	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.600%
62) 4-Bromofluorobenzene	11.079	95	56335	50.789	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	101.580%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	29032	63.157	ug/l	95
18) Methyl Acetate	2.709	43	2545	2.405	ug/l	93
20) Methylene Chloride	2.788	84	4334	5.036	ug/l	83
25) 2-Butanone	4.574	43	3880	5.766	ug/l	96
43) Isopropyl Acetate	6.342	43	9235	4.105	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040825\
Data File : VX045655.D
Acq On : 08 Apr 2025 16:21
Operator : JC/MD
Sample : Q1745-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IB-6A-WC

Quant Time: Apr 09 01:41:21 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

