

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040825\
 Data File : VX045658.D
 Acq On : 08 Apr 2025 17:31
 Operator : JC/MD
 Sample : Q1745-12
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IB-6.5-WC

Quant Time: Apr 09 01:42:48 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	63634	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	131134	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	125823	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	54135	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	67949	58.390	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	116.780%
35) Dibromofluoromethane	5.385	113	48605	52.241	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.480%
50) Toluene-d8	8.647	98	167623	51.617	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	103.240%
62) 4-Bromofluorobenzene	11.079	95	63055	53.306	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	106.620%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	27864	58.311	ug/l	100
18) Methyl Acetate	2.709	43	2342	2.129	ug/l #	86
20) Methylene Chloride	2.782	84	4117	4.602	ug/l	98
25) 2-Butanone	4.593	43	4198	6.002	ug/l	95
43) Isopropyl Acetate	6.348	43	9595	3.999	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040825\
Data File : VX045658.D
Acq On : 08 Apr 2025 17:31
Operator : JC/MD
Sample : Q1745-12
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IB-6.5-WC

Quant Time: Apr 09 01:42:48 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

