

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040825\
 Data File : VX045659.D
 Acq On : 08 Apr 2025 17:54
 Operator : JC/MD
 Sample : Q1743-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP-16

Quant Time: Apr 09 01:43:16 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	61974	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	128932	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	119271	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	50866	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	65097	57.438	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	114.880%
35) Dibromofluoromethane	5.385	113	46589	50.930	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.860%
50) Toluene-d8	8.647	98	163095	51.080	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	102.160%
62) 4-Bromofluorobenzene	11.079	95	59663	51.300	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	102.600%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	101259	217.582	ug/l	99
18) Methyl Acetate	2.709	43	2160	2.016	ug/l #	90
20) Methylene Chloride	2.788	84	10515	12.069	ug/l	97
25) 2-Butanone	4.568	43	11599	17.027	ug/l #	90
43) Isopropyl Acetate	6.348	43	7832	3.320	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040825\
Data File : VX045659.D
Acq On : 08 Apr 2025 17:54
Operator : JC/MD
Sample : Q1743-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP-16

Quant Time: Apr 09 01:43:16 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

