

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040825\
 Data File : VX045661.D
 Acq On : 08 Apr 2025 18:40
 Operator : JC/MD
 Sample : Q1744-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 B-158-SB02

Quant Time: Apr 09 01:44:13 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	63944	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	127369	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	121646	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	51074	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	63919	54.661	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	109.320%
35) Dibromofluoromethane	5.379	113	46324	51.261	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.520%
50) Toluene-d8	8.647	98	163599	51.867	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	103.740%
62) 4-Bromofluorobenzene	11.079	95	60955	53.054	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	106.100%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	48261	100.507	ug/l	96
18) Methyl Acetate	2.709	43	2379	2.152	ug/l #	90
20) Methylene Chloride	2.782	84	4803	5.343	ug/l	97
25) 2-Butanone	4.574	43	10995	15.643	ug/l	93
43) Isopropyl Acetate	6.354	43	7289	3.128	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040825\
Data File : VX045661.D
Acq On : 08 Apr 2025 18:40
Operator : JC/MD
Sample : Q1744-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
B-158-SB02

Quant Time: Apr 09 01:44:13 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

