

Data File : VX008769.D
Acq On : 09 Apr 2019 14:38
Operator : JC/SP
Sample : K2306-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT-TB-20190404

Quant Time: Apr 09 16:51:43 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X040319W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 04 04:47:41 2019
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	196170	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	322192	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	280372	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	116367	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	133181	46.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.10%	
35) Dibromofluoromethane	5.49	113	97651	47.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.80%	
50) Toluene-d8	8.71	98	406460	49.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.10%	
62) 4-Bromofluorobenzene	11.13	95	133822	46.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.02%	
Target Compounds						
16) Acetone	2.44	43	9946	6.035	ug/l	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VX008769.D
Acq On : 09 Apr 2019 14:38
Operator : JC/SP
Sample : K2306-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT-TB-20190404

Quant Time: Apr 09 16:51:43 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X040319W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 04 04:47:41 2019
Response via : Initial Calibration

