

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX040919\
 Data File : VX008776.D
 Acq On : 09 Apr 2019 17:22
 Operator : JC/SP
 Sample : K2311-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 0236-COMP-CS-1-ELTRT

Quant Time: Apr 10 07:32:20 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X040319W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 04 04:47:41 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	207288	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	343402	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	293774	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	121082	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	143151	47.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.70%	
35) Dibromofluoromethane	5.50	113	104492	47.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.18%	
50) Toluene-d8	8.71	98	429532	49.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.26%	
62) 4-Bromofluorobenzene	11.13	95	139770	45.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.18%	
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	3.04	59	5882	8.310	ug/l	97
16) Acetone	2.44	43	61145	35.113	ug/l	96
25) 2-Butanone	4.68	43	20837	7.571	ug/l	98
84) 1,2,4-Trimethylbenzene	11.80	105	3363	0.394	ug/l	94
95) Naphthalene	13.83	128	7877	0.791	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX040919\
Data File : VX008776.D
Acq On : 09 Apr 2019 17:22
Operator : JC/SP
Sample : K2311-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
0236-COMP-CS-1-ELTRT

Quant Time: Apr 10 07:32:20 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X040319W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 04 04:47:41 2019
Response via : Initial Calibration

