

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX040919\
 Data File : VX008777.D
 Acq On : 09 Apr 2019 17:45
 Operator : JC/SP
 Sample : K2311-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 0237-COMP-CS-2-ELTRT

Quant Time: Apr 10 07:37:24 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X040319W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 04 04:47:41 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	208982	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	344584	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	297056	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	123956	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	143058	46.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.86%	
35) Dibromofluoromethane	5.50	113	105777	48.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.02%	
50) Toluene-d8	8.71	98	430089	49.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.06%	
62) 4-Bromofluorobenzene	11.13	95	142227	45.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.44%	
Target Compounds						
11) Tert butyl alcohol	3.02	59	9240	12.948	ug/l #	87
16) Acetone	2.45	43	43796	24.946	ug/l	97
25) 2-Butanone	4.68	43	15338	5.528	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX040919\
Data File : VX008777.D
Acq On : 09 Apr 2019 17:45
Operator : JC/SP
Sample : K2311-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
0237-COMP-CS-2-ELTRT

Quant Time: Apr 10 07:37:24 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X040319W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 04 04:47:41 2019
Response via : Initial Calibration

