

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX040919\
 Data File : VX008780.D
 Acq On : 09 Apr 2019 18:55
 Operator : JC/SP
 Sample : K2311-07
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 0240-COMP-CS-5-ELTRT

Quant Time: Apr 10 07:54:25 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X040319W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 04 04:47:41 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	207625	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	343430	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	296307	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	125574	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	142800	47.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.30%	
35) Dibromofluoromethane	5.50	113	103565	47.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.34%	
50) Toluene-d8	8.71	98	428844	49.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.10%	
62) 4-Bromofluorobenzene	11.13	95	141696	45.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.40%	
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	3.03	59	7078	9.983	ug/l #	72
16) Acetone	2.44	43	55998	32.105	ug/l	96
25) 2-Butanone	4.68	43	17966	6.517	ug/l	97
79) 2-Chlorotoluene	11.42	91	5269	0.719	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX040919\
Data File : VX008780.D
Acq On : 09 Apr 2019 18:55
Operator : JC/SP
Sample : K2311-07
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
0240-COMP-CS-5-ELTRT

Quant Time: Apr 10 07:54:25 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X040319W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 04 04:47:41 2019
Response via : Initial Calibration

