

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX040919\
 Data File : VX008786.D
 Acq On : 09 Apr 2019 21:15
 Operator : JC/SP
 Sample : K2338-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP-1-SHALLOW

Quant Time: Apr 10 08:08:07 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X040319W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 04 04:47:41 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	209705	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	346559	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	305140	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	133509	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	142729	46.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.32%	
35) Dibromofluoromethane	5.50	113	103883	46.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.76%	
50) Toluene-d8	8.71	98	435715	49.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.78%	
62) 4-Bromofluorobenzene	11.13	95	147970	47.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.60%	
Target Compounds						
16) Acetone	2.44	43	15261	8.663	ug/l	100
20) Methylene Chloride	2.86	84	119030	40.364	ug/l	98
43) Isopropyl Acetate	6.45	43	12171	1.537	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX040919\
Data File : VX008786.D
Acq On : 09 Apr 2019 21:15
Operator : JC/SP
Sample : K2338-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP-1-SHALLOW

Quant Time: Apr 10 08:08:07 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X040319W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 04 04:47:41 2019
Response via : Initial Calibration

