

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX040919\
 Data File : VX008761.D
 Acq On : 09 Apr 2019 10:28
 Operator : JC/SP
 Sample : VX0409WBL01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0409WBL01

Quant Time: Apr 09 14:08:53 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA X\METHOD\624X032719W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Mar 27 12:01:15 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.01	128	34044	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	6.85	114	197596	30.00	ug/l	0.00
57) Chlorobenzene-d5	10.11	117	165238	30.00	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	6.06	65	84101	29.53	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	98.43%
60) 4-Bromofluorobenzene	11.13	95	77963	27.08	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	90.27%
63) Toluene-d8	8.71	98	237277	30.25	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	100.83%


Target Compounds	Ovalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX040919\
Data File : VX008761.D
Acq On : 09 Apr 2019 10:28
Operator : JC/SP
Sample : VX0409WBL01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0409WBL01

Quant Time: Apr 09 14:08:53 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA X\METHOD\624X032719W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Wed Mar 27 12:01:15 2019
Response via : Initial Calibration





#1

Bromochloromethane

Concen: 30.000 ug/l

RT: 5.01 min Scan# 633

Delta R.T. -0.01 min

Lab File: VX008761.D

Acq: 09 Apr 2019 10:28

Instrument :

MSVOA_X

ClientSampleId :

VX0409WBL01


Tot Ion:128 Resp: 34044

Ion Ratio Lower Upper

128 100

49 203.1 0.0 350.8




130 126.5 0.0 325.8



Abundance Ion 128.00 (127.70 to 128.70): V

Ion 49.00 (48.70 to 49.70): VX0

Ion 130.00 (129.70 to 130.70): V



#27

1,2-Dichloroethane-d4

Concen: 29.531 ug/l

RT: 6.06 min Scan# 805

Delta R.T. -0.00 min

Lab File: VX008761.D


Acq: 09 Apr 2019 10:28

Tgt Ion: 65 Resp: 84101

Ion Ratio Lower Upper



65 100


67 54.2 42.9 64.3



Abundance Ion 65.00 (64.70 to 65.70): VX0

Ion 67.00 (66.70 to 67.70): VX0









#28
 1,4-Difluorobenzene
 Concen: 30.000 ug/l
 RT: 6.85 min Scan# 935
 Delta R.T. -0.01 min
 Lab File: VX008761.D
 Acq: 09 Apr 2019 10:28




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0409WBL01


Tot Ion:114 Resp: 197596
 Ion Ratio Lower Upper
 114 100
 63 23.1 14.2 21.4#
 88 16.5 11.9 17.9



#57
 Chlorobenzene-d5
 Concen: 30.000 ug/l
 RT: 10.11 min Scan# 1469
 Delta R.T. -0.00 min
 Lab File: VX008761.D
 Acq: 09 Apr 2019 10:28

Tgt Ion:117 Resp: 165238
 Ion Ratio Lower Upper
 117 100
 82 59.2 40.4 60.6
 119 31.6 25.8 38.6









#60
4-Bromofluorobenzene
Concen: 27.080 ug/l
RT: 11.13 min Scan# 1637
Delta R.T. -0.00 min
Lab File: VX008761.D
Acq: 09 Apr 2019 10:28

Instrument :
MSVOA_X
ClientSampleId :
VX0409WBL01

Tot Ion: 95 Resp: 77963
Ion Ratio Lower Upper
95 100
174 76.9 74.9 112.3
176 75.3 70.7 106.1




Abundance Ion 95.00 (94.70 to 95.70): VX0
Ion 174.00 (173.70 to 174.70): V
Ion 176.00 (175.70 to 176.70): V



#63
Toluene-d8
Concen: 30.252 ug/l
RT: 8.71 min Scan# 1240
Delta R.T. -0.00 min
Lab File: VX008761.D
Acq: 09 Apr 2019 10:28

Tgt Ion: 98 Resp: 237277
Ion Ratio Lower Upper
98 100
100 63.5 50.9 76.3



Abundance Ion 98.00 (97.70 to 98.70): VX0
Ion 100.00 (99.70 to 100.70): V

