


Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX040919\
Data File : VX008759.D
Acq On : 09 Apr 2019 09:10
Operator : JC/SP
Sample : BFB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT3.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\624X032719W.M
Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
Last Update : Wed Mar 27 12:01:15 2019



AutoFind: Scans 1636, 1637, 1638; Background Corrected with Scan 1631

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.4	12686	PASS
75	95	30	60	51.6	33688	PASS
95	95	100	100	100.0	65242	PASS
96	95	5	9	6.7	4393	PASS
173	174	0.00	2	0.8	395	PASS
174	95	50	100	75.8	49461	PASS
175	174	5	9	7.9	3906	PASS
176	174	95	101	96.0	47464	PASS
177	176	5	9	6.5	3085	PASS