

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX040920\
 Data File : VX015589.D
 Acq On : 09 Apr 2020 12:22
 Operator : JC/SP
 Sample : L2135-08ME
 Misc : 12.9g/5.0mL/100uL/5.0mL/MSVOA_X/MEOH
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SB-3K-(0-2)ME

Quant Time: Apr 09 16:25:40 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 27 09:35:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	1009717	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	1569678	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	1465098	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	863641	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	630552	44.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.56%	
35) Dibromofluoromethane	5.47	113	474810	47.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.64%	
50) Toluene-d8	8.70	98	1881408	50.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.86%	
62) 4-Bromofluorobenzene	11.12	95	807979	53.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.64%	
Target Compounds						
31) Cyclohexane	5.54	56	38845	2.155	ug/l #	85
39) Methylcyclohexane	7.44	83	162852	9.766	ug/l	94
73) Isopropylbenzene	11.01	105	231200	4.086	ug/l	99
78) n-propylbenzene	11.35	91	567234	8.766	ug/l	100
85) sec-Butylbenzene	11.93	105	368525	6.750	ug/l	84
89) n-Butylbenzene	12.37	91	250622	5.472	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX040920\
Data File : VX015589.D
Acq On : 09 Apr 2020 12:22
Operator : JC/SP
Sample : L2135-08ME
Misc : 12.9g/5.0mL/100uL/5.0mL/MSVOA_X/MEOH
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SB-3K-(0-2)ME

Quant Time: Apr 09 16:25:40 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 27 09:35:51 2020
Response via : Initial Calibration

