

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX040920\
 Data File : VX015590.D
 Acq On : 09 Apr 2020 12:45
 Operator : JC/SP
 Sample : L2190-13DL 10X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ST008MW158-200407DL

Quant Time: Apr 10 06:13:34 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 27 09:35:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	1095208	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	1698125	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	1668002	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	922797	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	709312	45.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.84%	
35) Dibromofluoromethane	5.47	113	540436	49.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.58%	
50) Toluene-d8	8.70	98	2078743	52.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.02%	
62) 4-Bromofluorobenzene	11.12	95	884347	54.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.90%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.56	96	136676	10.295	ug/l	91
44) Trichloroethene	7.19	130	309337	24.748	ug/l	92
64) Tetrachloroethene	9.32	164	840051	66.734	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX040920\
Data File : VX015590.D
Acq On : 09 Apr 2020 12:45
Operator : JC/SP
Sample : L2190-13DL 10X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STOO8MW158-200407DL

Quant Time: Apr 10 06:13:34 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 27 09:35:51 2020
Response via : Initial Calibration

