

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX040920\
 Data File : VX015592.D
 Acq On : 09 Apr 2020 13:30
 Operator : JC/SP
 Sample : L2190-04 20X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 KY062MW128-200406

Quant Time: Apr 10 06:29:24 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 27 09:35:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	1063907	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	1695648	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	1614148	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	895286	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	700511	46.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.36%	
35) Dibromofluoromethane	5.47	113	525483	48.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.96%	
50) Toluene-d8	8.70	98	2025932	50.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.54%	
62) 4-Bromofluorobenzene	11.12	95	880637	54.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.60%	
Target Compounds						
4) Vinyl Chloride	1.39	62	31034	2.831	ug/l	94
21) trans-1,2-Dichloroethene	3.14	96	13304	1.193	ug/l	97
27) cis-1,2-Dichloroethene	4.57	96	407648	31.608	ug/l	91
44) Trichloroethene	7.19	130	1135764	90.996	ug/l	98
64) Tetrachloroethene	9.32	164	158201	12.987	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX040920\
Data File : VX015592.D
Acq On : 09 Apr 2020 13:30
Operator : JC/SP
Sample : L2190-04 20X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
KY062MW128-200406

Quant Time: Apr 10 06:29:24 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 27 09:35:51 2020
Response via : Initial Calibration

