

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX040920\
 Data File : VX015593.D
 Acq On : 09 Apr 2020 13:53
 Operator : JC/SP
 Sample : L2190-06 20X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 KY062MW139-200407

Quant Time: Apr 10 06:34:13 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 27 09:35:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	1074833	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	1672601	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	1580742	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	888229	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	703348	46.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.80%	
35) Dibromofluoromethane	5.47	113	513922	48.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.14%	
50) Toluene-d8	8.70	98	2007843	51.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.02%	
62) 4-Bromofluorobenzene	11.12	95	848392	53.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.06%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.58	96	14102	1.082	ug/l	83
44) Trichloroethene	7.19	130	9584	0.778	ug/l	97
64) Tetrachloroethene	9.32	164	59990	5.029	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX040920\
Data File : VX015593.D
Acq On : 09 Apr 2020 13:53
Operator : JC/SP
Sample : L2190-06 20X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
KY062MW139-200407

Quant Time: Apr 10 06:34:13 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 27 09:35:51 2020
Response via : Initial Calibration

