

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX040920\
 Data File : VX015594.D
 Acq On : 09 Apr 2020 14:16
 Operator : JC/SP
 Sample : L2190-08 10X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 KY062MW146-200407

Quant Time: Apr 10 06:52:12 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 27 09:35:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	1071015	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	1698643	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	1604935	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	894568	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	709369	46.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.92%	
35) Dibromofluoromethane	5.47	113	528759	48.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.40%	
50) Toluene-d8	8.70	98	2029324	50.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.52%	
62) 4-Bromofluorobenzene	11.12	95	871716	53.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.30%	
Target Compounds						
4) Vinyl Chloride	1.39	62	14663	1.329	ug/l	100
27) cis-1,2-Dichloroethene	4.57	96	243513	18.756	ug/l	92
44) Trichloroethene	7.20	130	102613	8.207	ug/l	91
64) Tetrachloroethene	9.32	164	782231	64.583	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX040920\
Data File : VX015594.D
Acq On : 09 Apr 2020 14:16
Operator : JC/SP
Sample : L2190-08 10X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
KY062MW146-200407

Quant Time: Apr 10 06:52:12 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 27 09:35:51 2020
Response via : Initial Calibration

