

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX040920\
 Data File : VX015596.D
 Acq On : 09 Apr 2020 15:01
 Operator : JC/SP
 Sample : L2190-17
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 M-16-200406

Quant Time: Apr 10 07:18:03 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 27 09:35:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	1070343	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	1652558	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	1578905	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	878372	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	694255	45.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.98%	
35) Dibromofluoromethane	5.47	113	518249	49.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.12%	
50) Toluene-d8	8.70	98	1994496	51.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.56%	
62) 4-Bromofluorobenzene	11.12	95	863021	54.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.20%	
Target Compounds						
16) Acetone	2.42	43	119271	18.390	ug/l	99
27) cis-1,2-Dichloroethene	4.57	96	11201	0.863	ug/l	81
44) Trichloroethene	7.20	130	15833	1.302	ug/l	84
64) Tetrachloroethene	9.32	164	49369	4.143	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX040920\
Data File : VX015596.D
Acq On : 09 Apr 2020 15:01
Operator : JC/SP
Sample : L2190-17
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampled :
M-16-200406

Quant Time: Apr 10 07:18:03 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 27 09:35:51 2020
Response via : Initial Calibration

