

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX040920\
 Data File : VX015600.D
 Acq On : 09 Apr 2020 16:32
 Operator : JC/SP
 Sample : L2190-24
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 Z3-DUP-0541-200406

Quant Time: Apr 10 08:47:45 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 27 09:35:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	1095303	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	1722389	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	1645951	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	927234	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	692923	44.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.72%	
35) Dibromofluoromethane	5.47	113	536300	48.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.42%	
50) Toluene-d8	8.70	98	2077324	51.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.50%	
62) 4-Bromofluorobenzene	11.12	95	886360	53.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.60%	
Target Compounds						
16) Acetone	2.44	43	131466	19.809	ug/l	99
27) cis-1,2-Dichloroethene	4.58	96	11344	0.854	ug/l	80
44) Trichloroethene	7.19	130	16852	1.329	ug/l	78
64) Tetrachloroethene	9.32	164	49565	3.990	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX040920\
Data File : VX015600.D
Acq On : 09 Apr 2020 16:32
Operator : JC/SP
Sample : L2190-24
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
Z3-DUP-0541-200406

Quant Time: Apr 10 08:47:45 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 27 09:35:51 2020
Response via : Initial Calibration

