

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX040920\
 Data File : VX015603.D
 Acq On : 09 Apr 2020 17:52
 Operator : JC/SP
 Sample : L2190-15
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SS037MW301-200406

Quant Time: Apr 10 09:01:34 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 27 09:35:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	1099797	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	1743177	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	1702188	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	936653	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.04	65	720784	46.47	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	92.94%	
35) Dibromofluoromethane	5.47	113	537901	48.27	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	96.54%	
50) Toluene-d8	8.70	98	2106106	51.34	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	102.68%	
62) 4-Bromofluorobenzene	11.12	95	926705	55.58	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	111.16%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
4) Vinyl Chloride	1.39	62	1843224	162.681	ug/l	100
12) 1,1-Dichloroethene	2.36	96	13355	1.286	ug/l	97
16) Acetone	2.42	43	24120	3.619	ug/l	96
21) trans-1,2-Dichloroethene	3.14	96	5610	0.487	ug/l	95
24) 1,1-Dichloroethane	3.68	63	31389	1.387	ug/l	97
27) cis-1,2-Dichloroethene	4.57	96	3459194	259.466	ug/l	90
44) Trichloroethene	7.19	130	33368	2.601	ug/l	87
64) Tetrachloroethene	9.32	164	51347	3.997	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX040920\
Data File : VX015603.D
Acq On : 09 Apr 2020 17:52
Operator : JC/SP
Sample : L2190-15
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SS037MW301-200406

Quant Time: Apr 10 09:01:34 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 27 09:35:51 2020
Response via : Initial Calibration

