

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX040920\
 Data File : VX015604.D
 Acq On : 09 Apr 2020 18:15
 Operator : JC/SP
 Sample : L2190-16
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ST008MW167-200406

Quant Time: Apr 10 09:05:47 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 27 09:35:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	1076018	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	1689302	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	1619111	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	889959	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.04	65	714640	47.09	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	94.18%	
35) Dibromofluoromethane	5.47	113	526586	48.76	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	97.52%	
50) Toluene-d8	8.70	98	2013295	50.64	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	101.28%	
62) 4-Bromofluorobenzene	11.12	95	857396	53.07	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	106.14%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
4) Vinyl Chloride	1.39	62	3557	0.321	ug/l #	74
16) Acetone	2.42	43	28802	4.417	ug/l	94
21) trans-1,2-Dichloroethene	3.15	96	10207	0.905	ug/l	95
27) cis-1,2-Dichloroethene	4.57	96	1539740	118.044	ug/l	91
44) Trichloroethene	7.19	130	148099	11.910	ug/l	97
64) Tetrachloroethene	9.32	164	3441298	281.634	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX040920\
Data File : VX015604.D
Acq On : 09 Apr 2020 18:15
Operator : JC/SP
Sample : L2190-16
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ST008MW167-200406

Quant Time: Apr 10 09:05:47 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 27 09:35:51 2020
Response via : Initial Calibration

