

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040924\
 Data File : VX040951.D
 Acq On : 09 Apr 2024 13:44
 Operator : JC/MD
 Sample : P2031-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RSB-01-37-040824

Quant Time: Apr 10 00:37:01 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040124W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 02 05:40:55 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	45056	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	80824	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	81875	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	44609	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	40913	50.608	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	101.220%
35) Dibromofluoromethane	5.379	113	28125	49.129	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.260%
50) Toluene-d8	8.647	98	104850	50.686	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.380%
62) 4-Bromofluorobenzene	11.079	95	44866	54.311	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	108.620%
Target Compounds						
						Qvalue
16) Acetone	2.380	43	1039	3.310	ug/l	98
30) Chloroform	5.093	83	5382	4.925	ug/l	94
44) Trichloroethene	7.135	130	552	0.979	ug/l	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040924\
Data File : VX040951.D
Acq On : 09 Apr 2024 13:44
Operator : JC/MD
Sample : P2031-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RSB-01-37-040824

Quant Time: Apr 10 00:37:01 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040124W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 02 05:40:55 2024
Response via : Initial Calibration

