

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040924\
 Data File : VX040952.D
 Acq On : 09 Apr 2024 14:07
 Operator : JC/MD
 Sample : P2031-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RSB-01-37-040824-FD

Quant Time: Apr 10 00:37:24 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040124W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 02 05:40:55 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	40914	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	76338	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	76043	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	40091	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	37406	50.954	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 101.900%			
35) Dibromofluoromethane	5.385	113	26321	48.679	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 97.360%			
50) Toluene-d8	8.647	98	97085	49.690	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 99.380%			
62) 4-Bromofluorobenzene	11.079	95	40774	52.258	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 104.520%			
Target Compounds						
16) Acetone	2.386	43	1045	3.666	ug/l	96
30) Chloroform	5.093	83	4708	4.745	ug/l	92
44) Trichloroethene	7.123	130	499	0.937	ug/l	85
52) Toluene	8.720	92	483	0.359	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040924\
Data File : VX040952.D
Acq On : 09 Apr 2024 14:07
Operator : JC/MD
Sample : P2031-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RSB-01-37-040824-FD

Quant Time: Apr 10 00:37:24 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040124W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 02 05:40:55 2024
Response via : Initial Calibration

