

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX040924\
 Data File : VX040941.D
 Acq On : 09 Apr 2024 09:07
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 10 00:32:00 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040124W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 02 05:40:55 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	96	0.00
2 T	Dichlorodifluoromethane	50.000	43.969	12.1	84	0.00
3 P	Chloromethane	50.000	49.856	0.3	95	0.00
4 C	Vinyl Chloride	50.000	52.687	-5.4#	102	0.00
5 T	Bromomethane	50.000	58.907	-17.8	101	0.00
6 T	Chloroethane	50.000	51.142	-2.3	100	0.00
7 T	Trichlorofluoromethane	50.000	49.851	0.3	96	0.00
8 T	Diethyl Ether	50.000	48.768	2.5	97	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	47.712	4.6	93	0.00
10 T	Methyl Iodide	50.000	44.090	11.8	83	0.00
11 T	Tert butyl alcohol	250.000	234.570	6.2	90	0.00
12 CM	1,1-Dichloroethene	50.000	46.375	7.2#	92	0.00
13 T	Acrolein	250.000	251.321	-0.5	96	0.00
14 T	Allyl chloride	50.000	47.916	4.2	91	0.00
15 T	Acrylonitrile	250.000	231.203	7.5	88	0.00
16 T	Acetone	250.000	248.390	0.6	98	0.00
17 T	Carbon Disulfide	50.000	42.004	16.0	86	0.00
18 T	Methyl Acetate	50.000	44.586	10.8	85	0.00
19 T	Methyl tert-butyl Ether	50.000	47.098	5.8	92	0.00
20 T	Methylene Chloride	50.000	43.947	12.1	92	0.00
21 T	trans-1,2-Dichloroethene	50.000	44.650	10.7	92	0.00
22 T	Diisopropyl ether	50.000	49.244	1.5	95	0.00
23 T	Vinyl Acetate	250.000	247.389	1.0	93	0.00
24 P	1,1-Dichloroethane	50.000	47.090	5.8	92	0.00
25 T	2-Butanone	250.000	231.598	7.4	88	0.00
26 T	2,2-Dichloropropane	50.000	48.668	2.7	94	0.00
27 T	cis-1,2-Dichloroethene	50.000	47.133	5.7	93	0.00
28 T	Bromochloromethane	50.000	48.898	2.2	99	0.00
29 T	Tetrahydrofuran	250.000	234.366	6.3	89	0.00
30 C	Chloroform	50.000	49.117	1.8#	95	0.00
31 T	Cyclohexane	50.000	46.791	6.4	90	0.00
32 T	1,1,1-Trichloroethane	50.000	49.315	1.4	92	0.00
33 S	1,2-Dichloroethane-d4	50.000	48.418	3.2	98	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	94	0.00
35 S	Dibromofluoromethane	50.000	51.138	-2.3	97	0.00
36 T	1,1-Dichloropropene	50.000	47.797	4.4	92	0.00
37 T	Ethyl Acetate	50.000	46.627	6.7	88	0.00
38 T	Carbon Tetrachloride	50.000	50.532	-1.1	94	0.00
39 T	Methylcyclohexane	50.000	47.267	5.5	90	0.00
40 TM	Benzene	50.000	48.743	2.5	93	0.00
41 T	Methacrylonitrile	50.000	50.840	-1.7	92	-0.01
42 TM	1,2-Dichloroethane	50.000	49.524	1.0	93	0.00
43 T	Isopropyl Acetate	50.000	49.670	0.7	91	0.00
44 TM	Trichloroethene	50.000	48.549	2.9	93	0.00
45 C	1,2-Dichloropropane	50.000	51.147	-2.3#	94	0.00
46 T	Dibromomethane	50.000	50.876	-1.8	95	0.00
47 T	Bromodichloromethane	50.000	50.742	-1.5	96	0.00
48 T	Methyl methacrylate	50.000	50.866	-1.7	90	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040924\
 Data File : VX040941.D
 Acq On : 09 Apr 2024 09:07
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 10 00:32:00 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040124W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 02 05:40:55 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	912.061	8.8	75	0.00
50 S	Toluene-d8	50.000	50.392	-0.8	97	0.00
51 T	4-Methyl-2-Pentanone	250.000	250.894	-0.4	91	0.00
52 CM	Toluene	50.000	49.009	2.0#	92	0.00
53 T	t-1,3-Dichloropropene	50.000	48.581	2.8	92	0.00
54 T	cis-1,3-Dichloropropene	50.000	51.922	-3.8	94	0.00
55 T	1,1,2-Trichloroethane	50.000	50.021	-0.0	92	0.00
56 T	Ethyl methacrylate	50.000	50.595	-1.2	90	0.00
57 T	1,3-Dichloropropane	50.000	49.773	0.5	94	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	257.500	-3.0	96	0.00
59 T	2-Hexanone	250.000	247.741	0.9	90	0.00
60 T	Dibromochloromethane	50.000	52.995	-6.0	98	0.00
61 T	1,2-Dibromoethane	50.000	49.603	0.8	93	0.00
62 S	4-Bromofluorobenzene	50.000	50.686	-1.4	99	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	95	0.00
64 T	Tetrachloroethene	50.000	47.998	4.0	93	0.00
65 PM	Chlorobenzene	50.000	48.242	3.5	96	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	51.156	-2.3	96	0.00
67 C	Ethyl Benzene	50.000	48.897	2.2#	93	0.00
68 T	m/p-Xylenes	100.000	97.942	2.1	93	0.00
69 T	o-Xylene	50.000	50.417	-0.8	95	0.00
70 T	Styrene	50.000	50.807	-1.6	93	0.00
71 P	Bromoform	50.000	49.687	0.6	94	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	96	0.00
73 T	Isopropylbenzene	50.000	48.750	2.5	93	0.00
74 T	N-amyl acetate	50.000	49.746	0.5	92	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	48.309	3.4	92	0.00
76 T	1,2,3-Trichloropropane	50.000	47.214	5.6	94	0.00
77 T	Bromobenzene	50.000	49.869	0.3	93	0.00
78 T	n-propylbenzene	50.000	48.723	2.6	94	0.00
79 T	2-Chlorotoluene	50.000	47.854	4.3	96	0.00
80 T	1,3,5-Trimethylbenzene	50.000	50.202	-0.4	94	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	47.532	4.9	87	0.00
82 T	4-Chlorotoluene	50.000	48.979	2.0	95	0.00
83 T	tert-Butylbenzene	50.000	49.216	1.6	95	0.00
84 T	1,2,4-Trimethylbenzene	50.000	50.269	-0.5	94	0.00
85 T	sec-Butylbenzene	50.000	49.773	0.5	95	0.00
86 T	p-Isopropyltoluene	50.000	50.327	-0.7	94	0.00
87 T	1,3-Dichlorobenzene	50.000	48.132	3.7	96	0.00
88 T	1,4-Dichlorobenzene	50.000	46.388	7.2	95	0.00
89 T	n-Butylbenzene	50.000	50.275	-0.5	95	0.00
90 T	Hexachloroethane	50.000	50.670	-1.3	95	0.00
91 T	1,2-Dichlorobenzene	50.000	48.152	3.7	96	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	48.074	3.9	85	0.00
93 T	1,2,4-Trichlorobenzene	50.000	46.840	6.3	94	0.00
94 T	Hexachlorobutadiene	50.000	46.349	7.3	93	0.00
95 T	Naphthalene	50.000	48.532	2.9	91	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040924\
Data File : VX040941.D
Acq On : 09 Apr 2024 09:07
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Apr 10 00:32:00 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040124W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 02 05:40:55 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	48.864	2.3	95	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6