

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX040924\
 Data File : VX040953.D
 Acq On : 09 Apr 2024 16:29
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 10 00:37:47 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040124W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 02 05:40:55 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	105	-0.01
2 T	Dichlorodifluoromethane	50.000	45.164	9.7	94	0.00
3 P	Chloromethane	50.000	49.548	0.9	103	0.00
4 C	Vinyl Chloride	50.000	54.444	-8.9#	115	0.00
5 T	Bromomethane	50.000	53.629	-7.3	100	0.00
6 T	Chloroethane	50.000	53.675	-7.3	115	0.00
7 T	Trichlorofluoromethane	50.000	49.993	0.0	105	0.00
8 T	Diethyl Ether	50.000	50.834	-1.7	110	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	48.174	3.7	103	0.00
10 T	Methyl Iodide	50.000	45.781	8.4	94	0.00
11 T	Tert butyl alcohol	250.000	247.025	1.2	103	0.00
12 CM	1,1-Dichloroethene	50.000	47.176	5.6#	101	0.00
13 T	Acrolein	250.000	252.328	-0.9	105	0.00
14 T	Allyl chloride	50.000	50.000	0.0	104	0.00
15 T	Acrylonitrile	250.000	242.812	2.9	100	0.00
16 T	Acetone	250.000	226.075	9.6	98	0.00
17 T	Carbon Disulfide	50.000	41.641	16.7	93	0.00
18 T	Methyl Acetate	50.000	47.176	5.6	98	0.00
19 T	Methyl tert-butyl Ether	50.000	49.880	0.2	106	0.00
20 T	Methylene Chloride	50.000	45.936	8.1	105	0.00
21 T	trans-1,2-Dichloroethene	50.000	46.056	7.9	104	0.00
22 T	Diisopropyl ether	50.000	51.101	-2.2	108	0.00
23 T	Vinyl Acetate	250.000	259.056	-3.6	106	0.00
24 P	1,1-Dichloroethane	50.000	48.886	2.2	104	0.00
25 T	2-Butanone	250.000	240.497	3.8	100	0.00
26 T	2,2-Dichloropropane	50.000	50.400	-0.8	106	0.00
27 T	cis-1,2-Dichloroethene	50.000	48.680	2.6	104	0.00
28 T	Bromochloromethane	50.000	48.210	3.6	106	0.00
29 T	Tetrahydrofuran	250.000	241.326	3.5	100	0.00
30 C	Chloroform	50.000	50.542	-1.1#	107	-0.01
31 T	Cyclohexane	50.000	47.270	5.5	99	0.00
32 T	1,1,1-Trichloroethane	50.000	50.704	-1.4	103	0.00
33 S	1,2-Dichloroethane-d4	50.000	51.848	-3.7	114	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	106	-0.01
35 S	Dibromofluoromethane	50.000	53.330	-6.7	114	-0.01
36 T	1,1-Dichloropropene	50.000	47.142	5.7	103	0.00
37 T	Ethyl Acetate	50.000	49.060	1.9	104	0.00
38 T	Carbon Tetrachloride	50.000	49.669	0.7	105	0.00
39 T	Methylcyclohexane	50.000	43.942	12.1	95	0.00
40 TM	Benzene	50.000	48.477	3.0	104	0.00
41 T	Methacrylonitrile	50.000	52.980	-6.0	108	-0.02
42 TM	1,2-Dichloroethane	50.000	50.366	-0.7	106	-0.01
43 T	Isopropyl Acetate	50.000	50.032	-0.1	104	-0.01
44 TM	Trichloroethene	50.000	48.464	3.1	104	0.00
45 C	1,2-Dichloropropane	50.000	51.462	-2.9#	107	0.00
46 T	Dibromomethane	50.000	50.349	-0.7	106	0.00
47 T	Bromodichloromethane	50.000	50.250	-0.5	107	0.00
48 T	Methyl methacrylate	50.000	52.127	-4.3	104	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040924\
 Data File : VX040953.D
 Acq On : 09 Apr 2024 16:29
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 10 00:37:47 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040124W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 02 05:40:55 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	877.470	12.3	81	0.00
50 S	Toluene-d8	50.000	51.339	-2.7	112	0.00
51 T	4-Methyl-2-Pentanone	250.000	252.846	-1.1	103	0.00
52 CM	Toluene	50.000	49.186	1.6#	104	0.00
53 T	t-1,3-Dichloropropene	50.000	49.202	1.6	106	0.00
54 T	cis-1,3-Dichloropropene	50.000	51.644	-3.3	106	0.00
55 T	1,1,2-Trichloroethane	50.000	50.349	-0.7	104	0.00
56 T	Ethyl methacrylate	50.000	51.864	-3.7	104	0.00
57 T	1,3-Dichloropropane	50.000	50.337	-0.7	108	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	252.523	-1.0	106	0.00
59 T	2-Hexanone	250.000	244.399	2.2	101	0.00
60 T	Dibromochloromethane	50.000	51.994	-4.0	108	0.00
61 T	1,2-Dibromoethane	50.000	50.600	-1.2	107	0.00
62 S	4-Bromofluorobenzene	50.000	53.769	-7.5	118	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	108	0.00
64 T	Tetrachloroethene	50.000	46.208	7.6	102	0.00
65 PM	Chlorobenzene	50.000	47.821	4.4	108	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	50.946	-1.9	108	0.00
67 C	Ethyl Benzene	50.000	48.691	2.6#	106	0.00
68 T	m/p-Xylenes	100.000	97.570	2.4	105	0.00
69 T	o-Xylene	50.000	49.504	1.0	106	0.00
70 T	Styrene	50.000	50.339	-0.7	105	0.00
71 P	Bromoform	50.000	48.812	2.4	105	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	104	0.00
73 T	Isopropylbenzene	50.000	50.845	-1.7	106	0.00
74 T	N-ethyl acetate	50.000	51.808	-3.6	105	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	49.141	1.7	103	0.00
76 T	1,2,3-Trichloropropane	50.000	50.484	-1.0	110	0.00
77 T	Bromobenzene	50.000	50.624	-1.2	103	0.00
78 T	n-propylbenzene	50.000	49.971	0.1	106	0.00
79 T	2-Chlorotoluene	50.000	50.167	-0.3	110	0.00
80 T	1,3,5-Trimethylbenzene	50.000	51.517	-3.0	106	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	49.646	0.7	100	0.00
82 T	4-Chlorotoluene	50.000	50.622	-1.2	107	0.00
83 T	tert-Butylbenzene	50.000	50.301	-0.6	105	0.00
84 T	1,2,4-Trimethylbenzene	50.000	50.525	-1.0	104	0.00
85 T	sec-Butylbenzene	50.000	48.826	2.3	101	0.00
86 T	p-Isopropyltoluene	50.000	49.082	1.8	100	0.00
87 T	1,3-Dichlorobenzene	50.000	47.938	4.1	104	0.00
88 T	1,4-Dichlorobenzene	50.000	46.318	7.4	103	0.00
89 T	n-Butylbenzene	50.000	47.165	5.7	97	0.00
90 T	Hexachloroethane	50.000	50.604	-1.2	104	0.00
91 T	1,2-Dichlorobenzene	50.000	46.904	6.2	102	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	48.767	2.5	94	0.00
93 T	1,2,4-Trichlorobenzene	50.000	44.647	10.7	98	0.00
94 T	Hexachlorobutadiene	50.000	38.497	23.0	85	0.00
95 T	Naphthalene	50.000	47.642	4.7	97	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040924\
Data File : VX040953.D
Acq On : 09 Apr 2024 16:29
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Apr 10 00:37:47 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040124W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 02 05:40:55 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	45.758	8.5	98	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6