



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040924\
Data File : VX040940.D
Acq On : 09 Apr 2024 08:37
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040124W.M
Title : SW846 8260
Last Update : Tue Apr 02 05:40:55 2024



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	22.7	4917	PASS
75	95	30	60	59.1	12782	PASS
95	95	100	100	100.0	21619	PASS
96	95	5	9	6.2	1343	PASS
173	174	0.00	2	0.8	131	PASS
174	95	50	100	72.8	15741	PASS
175	174	5	9	7.4	1161	PASS
176	174	95	101	95.0	14954	PASS
177	176	5	9	6.9	1032	PASS