

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040925\
 Data File : VX045667.D
 Acq On : 09 Apr 2025 11:49
 Operator : JC/MD
 Sample : VX0409WBS01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0409WBS01

Manual Integrations
 APPROVED

Reviewed By : John Carlone 04/10/2025
 Supervised By : Mahesh Dadoda 04/10/2025

Quant Time: Apr 10 01:33:49 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	94074	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	163361	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	141398	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	64175	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	92053	53.508	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 107.020%			
35) Dibromofluoromethane	5.379	113	63174	54.505	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 109.020%			
50) Toluene-d8	8.647	98	214655	53.060	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 106.120%			
62) 4-Bromofluorobenzene	11.079	95	77473	52.574	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 105.140%			
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	1.166	85	28296	20.113	ug/l	99
3) Chloromethane	1.301	50	27094	18.746	ug/l	96
4) Vinyl Chloride	1.374	62	24425	18.466	ug/l	98
5) Bromomethane	1.593	94	11395	18.169	ug/l	98
6) Chloroethane	1.672	64	14346	20.443	ug/l	100
7) Trichlorofluoromethane	1.880	101	38861	19.702	ug/l	96
8) Diethyl Ether	2.130	74	12161	18.364	ug/l	91
9) 1,1,2-Trichlorotrifluo...	2.319	101	23620	20.436	ug/l	95
10) Methyl Iodide	2.447	142	26699	18.520	ug/l	98
11) Tert butyl alcohol	2.965	59	22380	96.798	ug/l	97
12) 1,1-Dichloroethene	2.313	96	20770	18.391	ug/l	96
13) Acrolein	2.233	56	25659	80.591	ug/l	99
14) Allyl chloride	2.660	41	41579	19.404	ug/l	98
15) Acrylonitrile	3.062	53	68147	93.374	ug/l	99
16) Acetone	2.380	43	65896	93.280	ug/l	95
17) Carbon Disulfide	2.508	76	46029	16.502	ug/l	99
18) Methyl Acetate	2.703	43	32926	20.244	ug/l	100
19) Methyl tert-butyl Ether	3.111	73	76110	19.265	ug/l	100
20) Methylene Chloride	2.782	84	24229	18.320	ug/l	91
21) trans-1,2-Dichloroethene	3.087	96	21257	18.424	ug/l	97
22) Diisopropyl ether	3.757	45	79613	18.861	ug/l #	83
23) Vinyl Acetate	3.721	43	345218	94.659	ug/l	100
24) 1,1-Dichloroethane	3.605	63	45684	19.137	ug/l	99
25) 2-Butanone	4.556	43	97809	94.586	ug/l	98
26) 2,2-Dichloropropane	4.471	77	32813	21.275	ug/l	99
27) cis-1,2-Dichloroethene	4.489	96	26125	18.585	ug/l	97
28) Bromochloromethane	4.891	49	22802	19.766	ug/l	99
29) Tetrahydrofuran	5.007	42	62847	93.742	ug/l	99
30) Chloroform	5.086	83	47524	19.348	ug/l	98
31) Cyclohexane	5.458	56	39121	18.537	ug/l	95
32) 1,1,1-Trichloroethane	5.379	97	39802	19.150	ug/l	99
36) 1,1-Dichloropropene	5.684	75	30850	19.713	ug/l	98
37) Ethyl Acetate	4.715	43	35720	18.112	ug/l	99
38) Carbon Tetrachloride	5.666	117	33982	20.092	ug/l	96
39) Methylcyclohexane	7.373	83	37396	19.494	ug/l	95
40) Benzene	6.031	78	92251	19.301	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040925\
 Data File : VX045667.D
 Acq On : 09 Apr 2025 11:49
 Operator : JC/MD
 Sample : VX0409WBS01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0409WBS01

Quant Time: Apr 10 01:33:49 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By : John Carlone 04/10/2025
 Supervised By : Mahesh Dadoda 04/10/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.916	41	21224	20.397	ug/l	97
42) 1,2-Dichloroethane	6.086	62	40264	20.397	ug/l	99
43) Isopropyl Acetate	6.336	43	57645	19.285	ug/l	100
44) Trichloroethene	7.123	130	21805	19.194	ug/l	97
45) 1,2-Dichloropropane	7.428	63	23339	19.571	ug/l	100
46) Dibromomethane	7.580	93	17802	19.446	ug/l	98
47) Bromodichloromethane	7.818	83	35828	19.598	ug/l	98
48) Methyl methacrylate	7.690	41	30321	19.647	ug/l	99
49) 1,4-Dioxane	7.659	88	11977	428.937	ug/l	97
51) 4-Methyl-2-Pentanone	8.568	43	200067	100.967	ug/l	99
52) Toluene	8.714	92	55913	19.333	ug/l	99
53) t-1,3-Dichloropropene	8.976	75	30206	18.684	ug/l	99
54) cis-1,3-Dichloropropene	8.360	75	34931	20.333	ug/l	98
55) 1,1,2-Trichloroethane	9.147	97	22964	19.926	ug/l	96
56) Ethyl methacrylate	9.116	69	35297	19.790	ug/l	99
57) 1,3-Dichloropropane	9.305	76	39607	19.736	ug/l	98
58) 2-Chloroethyl Vinyl ether	8.238	63	88591	98.189	ug/l	100
59) 2-Hexanone	9.427	43	148084	100.919	ug/l	99
60) Dibromochloromethane	9.519	129	24830	19.763	ug/l	98
61) 1,2-Dibromoethane	9.604	107	23143	19.837	ug/l	99
64) Tetrachloroethene	9.269	164	20900	20.935	ug/l	98
65) Chlorobenzene	10.073	112	61480	20.348	ug/l	96
66) 1,1,1,2-Tetrachloroethane	10.159	131	20527	19.568	ug/l	99
67) Ethyl Benzene	10.189	91	108358	20.024	ug/l	98
68) m/p-Xylenes	10.299	106	79752	40.521	ug/l	99
69) o-Xylene	10.640	106	39137	20.186	ug/l	99
70) Styrene	10.653	104	64826	20.258	ug/l	98
71) Bromoform	10.799	173	15275	19.502	ug/l #	100
73) Isopropylbenzene	10.957	105	104506	20.330	ug/l	99
74) N-amyl acetate	10.842	43	49023	19.937	ug/l	98
75) 1,1,2,2-Tetrachloroethane	11.207	83	35954	19.910	ug/l	99
76) 1,2,3-Trichloropropane	11.238	75	31590m	20.180	ug/l	
77) Bromobenzene	11.195	156	24141	20.325	ug/l	98
78) n-propylbenzene	11.299	91	120369	20.329	ug/l	100
79) 2-Chlorotoluene	11.360	91	74998	19.813	ug/l	100
80) 1,3,5-Trimethylbenzene	11.451	105	88053	20.715	ug/l	99
81) trans-1,4-Dichloro-2-b...	11.018	75	8674	18.806	ug/l	95
82) 4-Chlorotoluene	11.451	91	86832	20.575	ug/l	100
83) tert-Butylbenzene	11.713	119	84807	20.152	ug/l	99
84) 1,2,4-Trimethylbenzene	11.750	105	86160	20.195	ug/l	99
85) sec-Butylbenzene	11.890	105	106901	20.682	ug/l	100
86) p-Isopropyltoluene	12.006	119	86083	20.201	ug/l	99
87) 1,3-Dichlorobenzene	11.969	146	42896	19.841	ug/l	99
88) 1,4-Dichlorobenzene	12.036	146	43700	19.938	ug/l	97
89) n-Butylbenzene	12.329	91	74429	20.141	ug/l	99
90) Hexachloroethane	12.536	117	14599	19.934	ug/l	98
91) 1,2-Dichlorobenzene	12.335	146	44205	20.565	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	12.939	75	7722	20.447	ug/l	96
93) 1,2,4-Trichlorobenzene	13.585	180	22901	19.140	ug/l	99
94) Hexachlorobutadiene	13.725	225	10679	20.712	ug/l	98
95) Naphthalene	13.774	128	87230	19.591	ug/l	99
96) 1,2,3-Trichlorobenzene	13.957	180	24734	19.742	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040925\
Data File : VX045667.D
Acq On : 09 Apr 2025 11:49
Operator : JC/MD
Sample : VX0409WBS01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0409WBS01

Manual Integrations
APPROVED

Reviewed By :John Carlone 04/10/2025
Supervised By :Mahesh Dadoda 04/10/2025

Quant Time: Apr 10 01:33:49 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

