

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040925\
 Data File : VX045677.D
 Acq On : 09 Apr 2025 15:43
 Operator : JC/MD
 Sample : PB167492TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB167492TB

Quant Time: Apr 10 01:36:01 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	66565	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	132585	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	125619	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	49122	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	67635	55.561	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	111.120%
35) Dibromofluoromethane	5.385	113	48016	51.043	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.080%
50) Toluene-d8	8.647	98	168107	51.199	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	102.400%
62) 4-Bromofluorobenzene	11.079	95	62629	52.367	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	104.740%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	52130	104.290	ug/l	96
18) Methyl Acetate	2.709	43	2101	1.826	ug/l #	86
20) Methylene Chloride	2.782	84	11787	12.596	ug/l	95
25) 2-Butanone	4.581	43	6831	9.336	ug/l	94
43) Isopropyl Acetate	6.348	43	8196	3.378	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040925\
Data File : VX045677.D
Acq On : 09 Apr 2025 15:43
Operator : JC/MD
Sample : PB167492TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB167492TB

Quant Time: Apr 10 01:36:01 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

