

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040925\
 Data File : VX045679.D
 Acq On : 09 Apr 2025 16:29
 Operator : JC/MD
 Sample : Q1746-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 B-149-SB02

Quant Time: Apr 10 01:36:25 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	66411	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	132306	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	127098	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	54289	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	68181	56.140	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	112.280%
35) Dibromofluoromethane	5.385	113	48876	52.067	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.140%
50) Toluene-d8	8.647	98	170273	51.968	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	103.940%
62) 4-Bromofluorobenzene	11.079	95	63846	53.497	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	107.000%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	40298	80.806	ug/l	97
18) Methyl Acetate	2.703	43	1949	1.697	ug/l	93
20) Methylene Chloride	2.782	84	4301	4.607	ug/l #	90
25) 2-Butanone	4.574	43	8961	12.275	ug/l	100
43) Isopropyl Acetate	6.354	43	6357	2.626	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040925\
Data File : VX045679.D
Acq On : 09 Apr 2025 16:29
Operator : JC/MD
Sample : Q1746-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
B-149-SB02

Quant Time: Apr 10 01:36:25 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

