

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX041019\
 Data File : VX008810.D
 Acq On : 10 Apr 2019 17:06
 Operator : JC/SP
 Sample : K2349-19
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ST006MW008-190409

Quant Time: Apr 11 07:53:16 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X040319W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 04 04:47:41 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	203455	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	335567	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	292254	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	124877	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	137645	46.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.76%	
35) Dibromofluoromethane	5.50	113	100957	47.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.12%	
50) Toluene-d8	8.71	98	420167	49.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.36%	
62) 4-Bromofluorobenzene	11.13	95	141762	46.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.60%	
Target Compounds						
4) Vinyl Chloride	1.41	62	6036	1.907	ug/l	100
27) cis-1,2-Dichloroethene	4.60	96	23069	7.887	ug/l	89
44) Trichloroethene	7.21	130	769	0.302	ug/l	95
64) Tetrachloroethene	9.34	164	2540	1.173	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX041019\
Data File : VX008810.D
Acq On : 10 Apr 2019 17:06
Operator : JC/SP
Sample : K2349-19
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ST006MW008-190409

Quant Time: Apr 11 07:53:16 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X040319W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 04 04:47:41 2019
Response via : Initial Calibration

