

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX041019\
 Data File : VX008811.D
 Acq On : 10 Apr 2019 17:29
 Operator : JC/SP
 Sample : K2349-20
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP-0507-190409

Quant Time: Apr 11 07:55:00 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X040319W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 04 04:47:41 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	202051	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	326164	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	279602	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	112285	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	134699	45.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.42%	
35) Dibromofluoromethane	5.50	113	99305	47.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.24%	
50) Toluene-d8	8.71	98	407511	49.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.16%	
62) 4-Bromofluorobenzene	11.13	95	131658	44.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.42%	
Target Compounds						
4) Vinyl Chloride	1.40	62	67196	21.373	ug/l	97
12) 1,1-Dichloroethene	2.37	96	1676	0.718	ug/l	87
21) trans-1,2-Dichloroethene	3.17	96	1130	0.447	ug/l	91
27) cis-1,2-Dichloroethene	4.60	96	313194	107.814	ug/l	90
44) Trichloroethene	7.21	130	9052	3.659	ug/l	94
64) Tetrachloroethene	9.34	164	6331	3.056	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX041019\
Data File : VX008811.D
Acq On : 10 Apr 2019 17:29
Operator : JC/SP
Sample : K2349-20
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP-0507-190409

Quant Time: Apr 11 07:55:00 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X040319W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 04 04:47:41 2019
Response via : Initial Calibration

