

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX041019\
 Data File : VX008813.D
 Acq On : 10 Apr 2019 18:16
 Operator : JC/SP
 Sample : K2349-11 10X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 KY062MW139-190409

Quant Time: Apr 11 07:57:44 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X040319W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 04 04:47:41 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	206695	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	335569	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	290428	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	122111	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	136041	45.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.24%	
35) Dibromofluoromethane	5.50	113	100740	46.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.90%	
50) Toluene-d8	8.71	98	423210	49.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.08%	
62) 4-Bromofluorobenzene	11.13	95	138802	45.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.64%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.60	96	19094	6.425	ug/l	91
44) Trichloroethene	7.21	130	3382	1.329	ug/l	93
64) Tetrachloroethene	9.34	164	24936	11.589	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX041019\
Data File : VX008813.D
Acq On : 10 Apr 2019 18:16
Operator : JC/SP
Sample : K2349-11 10X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
KY062MW139-190409

Quant Time: Apr 11 07:57:44 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X040319W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 04 04:47:41 2019
Response via : Initial Calibration

