

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX041019\
 Data File : VX008821.D
 Acq On : 10 Apr 2019 22:08
 Operator : JC/SP
 Sample : VX0410WBL02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0410WBL02

Quant Time: Apr 11 08:09:20 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X040319W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 04 04:47:41 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	202693	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	329843	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	290089	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	126507	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	135490	45.83	ug/l	0.00
Spiked Amount	50.000		Recovery	=	91.66%	
35) Dibromofluoromethane	5.50	113	99677	47.27	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.54%	
50) Toluene-d8	8.71	98	417197	49.68	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.36%	
62) 4-Bromofluorobenzene	11.13	95	142381	47.82	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.64%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX041019\
Data File : VX008821.D
Acq On : 10 Apr 2019 22:08
Operator : JC/SP
Sample : VX0410WBL02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0410WBL02

Quant Time: Apr 11 08:09:20 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X040319W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 04 04:47:41 2019
Response via : Initial Calibration

