

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX041019\
 Data File : VX008843.D
 Acq On : 11 Apr 2019 06:57
 Operator : JC/SP
 Sample : K2346-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 52 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VNJ-229

Quant Time: Apr 11 09:09:15 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X040319W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 11 08:05:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	190140	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	306497	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	271586	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	120134	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	127130	45.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.68%	
35) Dibromofluoromethane	5.50	113	92498	47.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.40%	
50) Toluene-d8	8.71	98	389101	49.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.74%	
62) 4-Bromofluorobenzene	11.13	95	134342	48.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.10%	
Target Compounds						
16) Acetone	2.44	43	10825	6.777	ug/l	97
18) Methyl Acetate	2.78	43	5991	1.546	ug/l	96
20) Methylene Chloride	2.85	84	59013	22.071	ug/l	98
43) Isopropyl Acetate	6.45	43	10312	1.472	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX041019\
Data File : VX008843.D
Acq On : 11 Apr 2019 06:57
Operator : JC/SP
Sample : K2346-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 52 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VNJ-229

Quant Time: Apr 11 09:09:15 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X040319W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 11 08:05:56 2019
Response via : Initial Calibration

