

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX041020\
 Data File : VX015614.D
 Acq On : 10 Apr 2020 12:33
 Operator : JC/SP
 Sample : VX0410WBL02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0410WBL02

Quant Time: Apr 11 01:14:36 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA X\METHOD\624X032520W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Thu Mar 26 02:31:53 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	4.98	128	191918	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	6.84	114	1098813	30.00	ug/l	0.00
57) Chlorobenzene-d5	10.10	117	999210	30.00	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	6.03	65	477556	31.84	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	106.13%
60) 4-Bromofluorobenzene	11.12	95	507771	29.69	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	98.97%
63) Toluene-d8	8.70	98	1288619	30.35	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	101.17%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX041020\
Data File : VX015614.D
Acq On : 10 Apr 2020 12:33
Operator : JC/SP
Sample : VX0410WBL02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0410WBL02

Quant Time: Apr 11 01:14:36 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA X\METHOD\624X032520W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Thu Mar 26 02:31:53 2020
Response via : Initial Calibration

