

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX041020\
 Data File : VX015630.D
 Acq On : 10 Apr 2020 18:34
 Operator : JC/SP
 Sample : L2190-16DL 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ST008MW167-200406DL

Quant Time: Apr 13 05:37:37 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 27 09:35:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	1125595	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	1749908	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	1700499	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	937780	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	712457	44.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.76%	
35) Dibromofluoromethane	5.47	113	549204	49.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.20%	
50) Toluene-d8	8.70	98	2106919	51.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.32%	
62) 4-Bromofluorobenzene	11.12	95	901888	53.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.78%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.57	96	257719	18.888	ug/l	91
44) Trichloroethene	7.19	130	24128	1.873	ug/l	92
64) Tetrachloroethene	9.32	164	488059	38.031	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX041020\
Data File : VX015630.D
Acq On : 10 Apr 2020 18:34
Operator : JC/SP
Sample : L2190-16DL 5X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ST008MW167-200406DL

Quant Time: Apr 13 05:37:37 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 27 09:35:51 2020
Response via : Initial Calibration

