

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX041020\
 Data File : VX015638.D
 Acq On : 10 Apr 2020 21:36
 Operator : JC/SP
 Sample : L2196-08
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB180-GW-58-60

Quant Time: Apr 11 03:22:06 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 27 09:35:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	1079910	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	1660556	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	1591612	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	889542	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	671362	44.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.16%	
35) Dibromofluoromethane	5.47	113	500364	47.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.28%	
50) Toluene-d8	8.70	98	1962623	50.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.44%	
62) 4-Bromofluorobenzene	11.12	95	852497	53.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.36%	
Target Compounds						
16) Acetone	2.42	43	40452	6.182	ug/l	98
52) Toluene	8.77	92	21464	0.777	ug/l	94
68) m/p-Xylenes	10.34	106	11753	0.561	ug/l	95
69) o-Xylene	10.69	106	6818	0.329	ug/l	73

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX041020\
Data File : VX015638.D
Acq On : 10 Apr 2020 21:36
Operator : JC/SP
Sample : L2196-08
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB180-GW-58-60

Quant Time: Apr 11 03:22:06 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 27 09:35:51 2020
Response via : Initial Calibration

