

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041025\
 Data File : VX045691.D
 Acq On : 10 Apr 2025 12:17
 Operator : JC/MD
 Sample : Q1748-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IB-1.5-WC

Quant Time: Apr 11 01:54:37 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	64958	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	127970	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	120693	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	51059	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	63254	53.248	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	106.500%
35) Dibromofluoromethane	5.385	113	46047	50.715	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.440%
50) Toluene-d8	8.647	98	162553	51.293	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	102.580%
62) 4-Bromofluorobenzene	11.079	95	60763	52.639	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	105.280%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	27479	56.334	ug/l	100
18) Methyl Acetate	2.703	43	1356	1.207	ug/l	97
20) Methylene Chloride	2.788	84	5546	6.073	ug/l	92
25) 2-Butanone	4.581	43	3857	5.402	ug/l	97
43) Isopropyl Acetate	6.355	43	8264	3.529	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041025\
Data File : VX045691.D
Acq On : 10 Apr 2025 12:17
Operator : JC/MD
Sample : Q1748-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IB-1.5-WC

Quant Time: Apr 11 01:54:37 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

