

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041025\
 Data File : VX045694.D
 Acq On : 10 Apr 2025 13:27
 Operator : JC/MD
 Sample : Q1754-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP-1-CONCRETE

Quant Time: Apr 11 01:55:56 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	63597	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	126520	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	115785	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	46306	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	63408	54.520	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 109.040%		
35) Dibromofluoromethane	5.379	113	40826	45.480	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 90.960%		
50) Toluene-d8	8.647	98	155088	49.498	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 99.000%		
62) 4-Bromofluorobenzene	11.079	95	56816	49.783	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 99.560%		
Target Compounds						
						Qvalue
16) Acetone	2.380	43	35424	74.175	ug/l	96
18) Methyl Acetate	2.709	43	3972	3.612	ug/l	96
20) Methylene Chloride	2.782	84	3438	3.845	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041025\
Data File : VX045694.D
Acq On : 10 Apr 2025 13:27
Operator : JC/MD
Sample : Q1754-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP-1-CONCRETE

Quant Time: Apr 11 01:55:56 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

