

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041025\
 Data File : VX045697.D
 Acq On : 10 Apr 2025 14:36
 Operator : JC/MD
 Sample : PB167534ZHE#03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB167534ZHE#03

Quant Time: Apr 11 01:57:14 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	61049	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	122509	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	116892	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	50571	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	61420	55.015	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 110.020%	
35) Dibromofluoromethane	5.385	113	43883	50.487	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.980%	
50) Toluene-d8	8.647	98	154986	51.085	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 102.180%	
62) 4-Bromofluorobenzene	11.079	95	60068	54.356	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 108.720%	
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	2.983	59	1860	12.397	ug/l	98
16) Acetone	2.386	43	103576	225.933	ug/l	97
18) Methyl Acetate	2.709	43	2233	2.116	ug/l	97
20) Methylene Chloride	2.782	84	11271	13.133	ug/l	95
25) 2-Butanone	4.568	43	11208	16.702	ug/l	99
43) Isopropyl Acetate	6.342	43	8378	3.738	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041025\
Data File : VX045697.D
Acq On : 10 Apr 2025 14:36
Operator : JC/MD
Sample : PB167534ZHE#03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB167534ZHE#03

Quant Time: Apr 11 01:57:14 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

