

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041025\
 Data File : VX045699.D
 Acq On : 10 Apr 2025 15:23
 Operator : JC/MD
 Sample : PB167534ZHE#05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB167534ZHE#05

Quant Time: Apr 11 01:58:12 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	62866	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	126149	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	118204	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	49672	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	63158	54.936	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 109.880%	
35) Dibromofluoromethane	5.379	113	45126	50.419	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.840%	
50) Toluene-d8	8.647	98	159464	51.045	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 102.080%	
62) 4-Bromofluorobenzene	11.079	95	58774	51.650	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 103.300%	
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	2.983	59	1828	11.831	ug/l #	86
16) Acetone	2.386	43	107675	228.086	ug/l	100
18) Methyl Acetate	2.703	43	2341	2.154	ug/l #	84
20) Methylene Chloride	2.788	84	11089	12.547	ug/l	96
25) 2-Butanone	4.568	43	11024	15.953	ug/l	91
43) Isopropyl Acetate	6.342	43	8697	3.768	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041025\
Data File : VX045699.D
Acq On : 10 Apr 2025 15:23
Operator : JC/MD
Sample : PB167534ZHE#05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB167534ZHE#05

Quant Time: Apr 11 01:58:12 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

