

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041025\
 Data File : VX045700.D
 Acq On : 10 Apr 2025 15:46
 Operator : JC/MD
 Sample : PB167534ZHE#06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB167534ZHE#06

Quant Time: Apr 11 01:58:37 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	64406	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	127502	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	122158	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	53290	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	64930	55.127	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 110.260%	
35) Dibromofluoromethane	5.385	113	46356	51.243	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.480%	
50) Toluene-d8	8.647	98	161540	51.161	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 102.320%	
62) 4-Bromofluorobenzene	11.079	95	62450	54.299	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 108.600%	
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	2.977	59	1897	11.984	ug/l #	77
16) Acetone	2.380	43	104598	216.270	ug/l	100
18) Methyl Acetate	2.703	43	2248	2.019	ug/l	95
20) Methylene Chloride	2.782	84	10841	11.973	ug/l	96
25) 2-Butanone	4.568	43	11215	15.841	ug/l	96
43) Isopropyl Acetate	6.349	43	7801	3.344	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041025\
Data File : VX045700.D
Acq On : 10 Apr 2025 15:46
Operator : JC/MD
Sample : PB167534ZHE#06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB167534ZHE#06

Quant Time: Apr 11 01:58:37 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

