

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041025\
 Data File : VX045701.D
 Acq On : 10 Apr 2025 16:09
 Operator : JC/MD
 Sample : PB167534ZHE#07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB167534ZHE#07

Quant Time: Apr 11 01:59:02 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	66501	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	132525	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	129453	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	55187	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	67678	55.650	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	111.300%
35) Dibromofluoromethane	5.379	113	48235	51.299	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.600%
50) Toluene-d8	8.647	98	171721	52.324	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	104.640%
62) 4-Bromofluorobenzene	11.079	95	66017	55.224	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	110.440%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	23388	46.834	ug/l	98
18) Methyl Acetate	2.709	43	2313	2.012	ug/l	100
20) Methylene Chloride	2.782	84	3203	3.426	ug/l	99
43) Isopropyl Acetate	6.342	43	9555	3.940	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041025\
Data File : VX045701.D
Acq On : 10 Apr 2025 16:09
Operator : JC/MD
Sample : PB167534ZHE#07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB167534ZHE#07

Quant Time: Apr 11 01:59:02 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

