

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041025\
 Data File : VX045706.D
 Acq On : 10 Apr 2025 18:05
 Operator : JC/MD
 Sample : PB167534ZHE#12
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB167534ZHE#12

Quant Time: Apr 11 02:01:27 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	64585	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	129412	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	123469	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	52693	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	64779	54.847	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 109.700%		
35) Dibromofluoromethane	5.379	113	46821	50.993	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 101.980%		
50) Toluene-d8	8.647	98	165216	51.553	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 103.100%		
62) 4-Bromofluorobenzene	11.079	95	63244	54.177	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 108.360%		
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	23783	49.038	ug/l	96
18) Methyl Acetate	2.709	43	2216	1.985	ug/l	94
20) Methylene Chloride	2.782	84	3341	3.680	ug/l	91
43) Isopropyl Acetate	6.342	43	9031	3.814	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041025\
Data File : VX045706.D
Acq On : 10 Apr 2025 18:05
Operator : JC/MD
Sample : PB167534ZHE#12
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB167534ZHE#12

Quant Time: Apr 11 02:01:27 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

