

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041025\
 Data File : VX045707.D
 Acq On : 10 Apr 2025 18:28
 Operator : JC/MD
 Sample : PB167534ZHE#13
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB167534ZHE#13

Quant Time: Apr 11 02:01:56 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.549	168	63768	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.756	114	128556	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	122961	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	51067	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	64809	55.575	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	111.160%
35) Dibromofluoromethane	5.379	113	46427	50.901	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.800%
50) Toluene-d8	8.646	98	163340	51.307	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	102.620%
62) 4-Bromofluorobenzene	11.079	95	61524	53.055	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	106.100%
Target Compounds						
					Qvalue	
16) Acetone	2.379	43	18905	39.480	ug/l	92
18) Methyl Acetate	2.708	43	2292	2.079	ug/l	98
20) Methylene Chloride	2.788	84	2962	3.304	ug/l	94
43) Isopropyl Acetate	6.342	43	9484	4.032	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041025\
Data File : VX045707.D
Acq On : 10 Apr 2025 18:28
Operator : JC/MD
Sample : PB167534ZHE#13
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB167534ZHE#13

Quant Time: Apr 11 02:01:56 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

