

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041025\
 Data File : VX045709.D
 Acq On : 10 Apr 2025 19:15
 Operator : JC/MD
 Sample : Q1749-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP-14

Quant Time: Apr 11 02:02:49 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	63143	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	126552	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	121332	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	51288	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	66034	57.186	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	114.380%
35) Dibromofluoromethane	5.379	113	46417	51.696	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.400%
50) Toluene-d8	8.647	98	164578	52.514	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	105.020%
62) 4-Bromofluorobenzene	11.079	95	61661	54.015	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	108.040%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	29931	63.124	ug/l	96
18) Methyl Acetate	2.709	43	2502	2.292	ug/l	98
20) Methylene Chloride	2.782	84	6326	7.126	ug/l	97
25) 2-Butanone	4.581	43	4015	5.785	ug/l	97
43) Isopropyl Acetate	6.342	43	8668	3.743	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041025\
Data File : VX045709.D
Acq On : 10 Apr 2025 19:15
Operator : JC/MD
Sample : Q1749-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP-14

Quant Time: Apr 11 02:02:49 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

