

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041025\
 Data File : VX045720.D
 Acq On : 11 Apr 2025 06:42
 Operator : JC/MD
 Sample : Q1753-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WC-2

Quant Time: Apr 11 07:30:59 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	66493	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	132267	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	124072	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	53444	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	66284	54.511	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 109.020%	
35) Dibromofluoromethane	5.379	113	47348	50.454	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.900%	
50) Toluene-d8	8.647	98	167546	51.151	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 102.300%	
62) 4-Bromofluorobenzene	11.079	95	62563	52.437	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 104.880%	
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	19189	38.430	ug/l	96
18) Methyl Acetate	2.709	43	1976	1.719	ug/l	100
20) Methylene Chloride	2.782	84	3308	3.539	ug/l	91
43) Isopropyl Acetate	6.348	43	8933	3.691	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041025\
Data File : VX045720.D
Acq On : 11 Apr 2025 06:42
Operator : JC/MD
Sample : Q1753-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WC-2

Quant Time: Apr 11 07:30:59 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

