

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041025\
 Data File : VX045724.D
 Acq On : 11 Apr 2025 08:15
 Operator : JC/MD
 Sample : Q1752-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP-2

Quant Time: Apr 11 08:46:41 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	62957	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	125281	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	119024	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	49027	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	63150	54.850	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 109.700%	
35) Dibromofluoromethane	5.385	113	45481	51.167	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.340%	
50) Toluene-d8	8.647	98	160095	51.602	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 103.200%	
62) 4-Bromofluorobenzene	11.079	95	60544	53.575	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 107.140%	
Target Compounds						
					Qvalue	
16) Acetone	2.379	43	31171	65.934	ug/l	100
18) Methyl Acetate	2.709	43	2309	2.121	ug/l	95
20) Methylene Chloride	2.788	84	6413	7.246	ug/l	94
25) 2-Butanone	4.580	43	4179	6.039	ug/l	93
43) Isopropyl Acetate	6.348	43	8525	3.719	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041025\
Data File : VX045724.D
Acq On : 11 Apr 2025 08:15
Operator : JC/MD
Sample : Q1752-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 45 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP-2

Quant Time: Apr 11 08:46:41 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

